


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2149, Rack A, Plate 1, Well A05, 496445
[C83DB452F6FA4E68] May 20, 2026 12:14:45 Pressure = 66 bar (65 to 67)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.515	F 0.60	2.9%	28811	
2	0.712	F 0.83	13.7%	138345	Acetylated F peak
3	0.853	F 1.00	66.9%	675276	Consistent with F
4	1.217	A 1.01	16.5%	166570	A peak
Total Area:				1009002	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range