


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1982, Rack A, Plate 1, Well C01, 171090
[DC72B607A3F006EA] Nov 19, 2025 11:24:12 Pressure = 69 bar (68 to 70)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.388	F	0.45	1.6%	15370	
2	0.497	F	0.58	0.6%	5595	
3	0.712	F	0.83	10.4%	96971	Acetylated F peak
4	0.857	F	1.00	63.4%	591560	Consistent with F
5	1.245	A	1.02	23.9%	222849	A peak - REVIEW
Total Area: 932345						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range