


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2127, Rack A, Plate 1, Well F03, 489327-P
[DC261CDF979486C0] Apr 21, 2026 11:54:03 Pressure = 62 bar (62 to 64)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.418	F	0.48	2.8%	14008	
2	0.503	F	0.58	4.6%	23215	
3	0.715	F	0.82	18.9%	95506	
4	0.865	F	1.00	43.2%	218519	Consistent with F
5	1.017	F	1.17	6.7%	33771	3
6	1.247	A	1.01	23.9%	121236	A peak - REVIEW
Total Area: 506255						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range