


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1824, Rack A, Plate 1, Well F11, 455896
 [C921D0F5DAD506E2] Jun 04, 2025 13:06:38 Pressure = 57 bar (57 to 58)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.395	F 0.47	1.7%	21372	
2	0.492	F 0.58	1.1%	13246	
3	0.728	F 0.87	17.8%	222789	Acetylated F peak
4	0.838	F 1.00	63.7%	796438	Consistent with F
5	1.007	F 1.20	1.2%	15473	
6	1.227	A 1.01	14.3%	178263	A peak
7	1.382	A 1.14	0.2%	2218	
Total Area:				1249799	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range