


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1884, Rack A, Plate 1, Well A07, 457243  
 [89216315D42C713D] Aug 13, 2025 10:55:33 Pressure = 61 bar (60 to 62)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.408	F 0.49	1.0%	7900	
2	0.493	F 0.59	1.5%	12266	
3	0.698	F 0.83	10.9%	87086	Acetylated F peak
4	0.833	F 0.99	54.3%	434757	Consistent with F
5	0.973	F 1.16	3.4%	26878	
6	1.217	A 1.01	28.9%	231355	A peak - REVIEW
Total Area:				800242	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range