


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2054, Rack A, Plate 1, Well F02, 488995
[88629FDF97CA8648] Jan 26, 2026 12:47:13 Pressure = 87 bar (86 to 88)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.412	F 0.48	1.1%	10997	
2	0.497	F 0.58	1.4%	13383	
3	0.833	F 0.97	83.9%	802745	Consistent with F
4	1.198	A 1.01	13.6%	130037	A peak

Total Area: 957162

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range