


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2008, Rack A, Plate 1, Well D08, 171402
 [9D72FB3179C7F9B7] Dec 09, 2025 12:22:38 Pressure = 72 bar (72 to 74)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.495	F 0.57	5.1%	32414	
2	0.713	F 0.82	18.4%	117930	
3	0.875	F 1.01	51.9%	332062	Consistent with F
4	1.005	F 1.16	3.3%	21383	
5	1.107	A 0.91	0.2%	1485	
6	1.227	A 1.01	21.0%	134281	A peak - REVIEW
Total Area:				639555	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range