


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1993, Rack A, Plate 1, Well D01, 483304
[9937A992FDB884E0] Nov 25, 2025 11:47:59 Pressure = 68 bar (68 to 70)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.46	1.5%	13566	
2	0.493	F 0.56	0.5%	4115	
3	0.717	F 0.82	12.4%	109034	Acetylated F peak
4	0.860	F 0.98	63.7%	558034	Consistent with F
5	1.245	A 1.01	21.9%	191841	A peak - REVIEW

Total Area: 876590

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range