


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2146, Rack A, Plate 1, Well F10, 496444
[DC3DF190AAD06C68] May 13, 2026 14:04:53 Pressure = 66 bar (65 to 67)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.508	F 0.58	2.3%	23183	
2	0.695	F 0.80	10.0%	102412	Acetylated F peak
3	0.853	F 0.98	51.2%	526781	Consistent with F
4	1.230	A 1.01	36.6%	376857	A peak - REVIEW
Total Area:				1029233	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range