


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1833, Rack A, Plate 1, Well E03, 456665
 [C831C2EDE7FF8E40] Jun 18, 2025 11:42:43 Pressure = 59 bar (59 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.405	F 0.47	1.2%	14690	
2	0.495	F 0.58	1.2%	14542	
3	0.710	F 0.83	14.8%	185662	Acetylated F peak
4	0.840	F 0.98	65.8%	824469	Consistent with F
5	1.220	A 1.01	17.1%	213959	A peak
Total Area:				1253322	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range