


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1901, Rack A, Plate 1, Well A12, 830838
[D9743F6426992660] Sep 03, 2025 11:14:31 Pressure = 62 bar (61 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.48	2.0%	11120	
2	0.488	F 0.58	1.2%	6553	
3	0.672	F 0.80	12.3%	67856	Acetylated F peak
4	0.838	F 1.00	55.0%	304269	Consistent with F
5	0.978	F 1.16	3.7%	20199	
6	1.215	A 1.01	25.8%	142881	A peak - REVIEW
Total Area:				552878	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range