


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2086, Rack A, Plate 1, Well E09, 489597
 [99367971711B7335] Mar 04, 2026 12:48:23 Pressure = 53 bar (53 to 55)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.413	F 0.47	1.0%	14563	
2	0.515	F 0.58	0.9%	12686	
3	0.733	F 0.83	14.4%	205559	Acetylated F peak
4	0.872	F 0.99	58.2%	831858	Consistent with F
5	1.018	F 1.16	2.3%	32604	
6	1.268	A 1.01	23.3%	332901	A peak - REVIEW
Total Area:				1430171	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range