


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2029, Rack A, Plate 1, Well C06, 490080  
[D82782FDC55DF31D] Dec 29, 2025 11:52:49 Pressure = 74 bar (74 to 77)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.503	F 0.59	1.3%	20068	
2	0.708	F 0.83	15.8%	252619	Acetylated F peak
3	0.842	F 0.99	71.2%	1139395	Consistent with F
4	1.215	A 1.01	11.7%	187726	A peak
Total Area:				1599808	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range