


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2173, Rack A, Plate 1, Well A10, 486001
[CD69B92786C7BB1F] Jun 10, 2026 13:19:44

Pressure = 68 bar (67 to 69)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.418	F 0.50	1.3%	7328	
2	0.505	F 0.60	0.8%	4487	
3	0.687	F 0.82	8.6%	48577	Acetylated F peak
4	0.842	F 1.00	44.8%	251782	Consistent with F
5	1.193	A 1.01	44.5%	250005	A peak - REVIEW
Total Area:				562179	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range