


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2143, Rack A, Plate 1, Well D03, 496520
 [88771F09434BD91D] May 08, 2026 12:42:41 Pressure = 64 bar (63 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.515	F 0.60	1.2%	36923	
2	0.713	F 0.83	15.1%	450201	Acetylated F peak
3	0.840	F 0.98	64.0%	1914054	Consistent with F
4	1.215	A 1.01	19.7%	587248	A peak
Total Area:				2988426	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range