


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2130, Rack A, Plate 1, Well F10, 502217  
[C8774D7F9134AE62] Apr 24, 2026 13:34:46 Pressure = 62 bar (61 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.425	F 0.49	0.8%	9737	
2	0.513	F 0.60	0.8%	8958	
3	0.717	F 0.83	13.8%	162336	Acetylated F peak
4	0.850	F 0.99	63.6%	751126	Consistent with F
5	1.228	A 1.00	21.0%	248452	A peak - REVIEW

Total Area: 1180609

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range