


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1833, Rack A, Plate 1, Well E05, 785489
 [887597B8B2A87137] Jun 18, 2025 11:46:46 Pressure = 60 bar (59 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.395	F 0.46	1.2%	11942	
2	0.490	F 0.57	0.9%	9749	
3	0.725	F 0.85	12.4%	127611	Acetylated F peak
4	0.838	F 0.98	54.1%	555183	Consistent with F
5	1.000	F 1.17	0.8%	8015	
6	1.230	A 1.02	30.2%	310439	A peak - REVIEW
7	1.387	S 0.89	0.4%	3599	
Total Area:				1026538	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range