


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2133, Rack A, Plate 1, Well E02, 502926
[CC735AEC3A3EAC68] Apr 29, 2026 12:50:13 Pressure = 66 bar (65 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.513	F 0.60	2.5%	25198	
2	0.713	F 0.83	17.5%	177239	Acetylated F peak
3	0.853	F 0.99	66.8%	675188	Consistent with F
4	1.227	A 1.00	13.2%	132901	A peak

Total Area: 1010526

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range