


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2146, Rack A, Plate 1, Well E06, 496541
[9C7DF0C5FF053B35] May 13, 2026 13:30:31 Pressure = 66 bar (64 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.513	F 0.60	2.1%	21780	
2	0.710	F 0.83	13.1%	133132	Acetylated F peak
3	0.865	F 1.01	58.7%	596035	Consistent with F
4	1.222	A 1.00	26.0%	264097	A peak - REVIEW
Total Area:				1015044	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range