


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2166, Rack A, Plate 1, Well E01, 501539  
[DC28B71E4B70C640] Jun 02, 2026 11:41:07 Pressure = 63 bar (63 to 65)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.413	F 0.49	0.9%	8429	
2	0.508	F 0.61	0.5%	4665	
3	0.692	F 0.83	12.9%	114936	Acetylated F peak
4	0.847	F 1.01	63.8%	568024	Consistent with F
5	1.193	A 1.01	21.8%	193861	A peak - REVIEW
Total Area:				889915	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range