


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2054, Rack A, Plate 1, Well D07, 489166
[9D778F9F9DE00660] Jan 26, 2026 12:06:47 Pressure = 87 bar (86 to 89)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.503	F	0.60	1.5%	18085	
2	0.697	F	0.83	13.4%	158130	Acetylated F peak
3	0.833	F	0.99	71.4%	845005	Consistent with F
4	1.190	A	1.00	13.8%	162950	A peak
Total Area: 1184170						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range