


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2146, Rack A, Plate 1, Well E05, 496540
[887CF0C5FF07BBB7] May 13, 2026 13:28:30 Pressure = 66 bar (65 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.515	F 0.60	2.0%	22550	
2	0.710	F 0.83	13.5%	149647	Acetylated F peak
3	0.853	F 1.00	72.6%	803127	Consistent with F
4	1.230	A 1.01	11.9%	131571	A peak
Total Area:				1106895	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range