


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2098, Rack A, Plate 1, Well B03, 172736
 [CD362B0156317315] Mar 19, 2026 11:42:11 Pressure = 60 bar (60 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.418	F 0.48	3.3%	17008	
2	0.508	F 0.59	2.6%	13453	
3	0.725	F 0.83	17.8%	91896	
4	0.877	F 1.01	44.3%	228372	Consistent with F
5	0.997	F 1.15	4.6%	23978	
6	1.100	A 0.90	0.3%	1731	
7	1.243	A 1.01	27.0%	139315	A peak - REVIEW
Total Area:				515753	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range