


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2197, Rack A, Plate 1, Well B02, 487152
[996C92CAEA22C6E2] Jul 15, 2026 12:43:10 Pressure = 83 bar (83 to 85)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.515	F	0.62	1.0%	10131	
2	0.683	F	0.82	15.4%	151713	Acetylated F peak
3	0.820	F	0.99	73.6%	723306	Consistent with F
4	1.160	A	1.00	9.9%	97169	A peak
Total Area: 982319						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range