


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1380, Rack A, Plate 1, Well B07, 27414
[C89DF31BB40715BD] Apr 17, 2024 11:01:01

Pressure = 83 bar (83 to 84)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.395	F 0.47	1.5%	18916	
2	0.478	F 0.56	1.4%	17728	
3	0.755	F 0.89	14.7%	187594	Acetylated F peak
4	0.867	F 1.02	65.1%	831390	Consistent with F
5	1.218	A 1.01	17.3%	220830	A peak
Total Area:				1276458	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range