


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1865, Rack A, Plate 1, Well B11, 169132
 [8C208A8FBFE8D99F] Jul 16, 2025 11:04:40 Pressure = 59 bar (59 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.397	F 0.47	1.8%	12800	
2	0.492	F 0.59	1.5%	10737	
3	0.708	F 0.84	10.0%	70883	Acetylated F peak
4	0.858	F 1.02	42.7%	302352	Consistent with F
5	0.978	F 1.16	3.0%	21616	
6	1.103	A 0.91	1.0%	7287	
7	1.217	A 1.01	39.9%	283087	A peak - REVIEW
Total Area:				708762	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range