


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2114, Rack A, Plate 1, Well H10, 502305 FA
 [CC3346C275D6A440] Apr 07, 2026 16:45:23 Pressure = 59 bar (57 to 60)



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|------|-------|--------|--------|---------|-------------------|
| 1 | 0.417 | F 0.21 | 1.5% | 46851 | |
| 2 | 0.572 | F 0.29 | 2.5% | 77897 | |
| 3 | 0.923 | F 0.47 | 17.6% | 557952 | |
| 4 | 1.900 | F 0.97 | 53.0% | 1681989 | Consistent with F |
| 5 | 2.675 | F 1.37 | 1.8% | 55649 | |
| 6 | 2.958 | A 0.79 | 0.4% | 11351 | |
| 7 | 3.237 | A 0.86 | 0.2% | 5714 | |
| 8 | 3.790 | A 1.01 | 22.9% | 725854 | A peak - REVIEW |
| 9 | 4.490 | A 1.19 | 0.4% | 11255 | |

Total Area: 3174512

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range