


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1842, Rack A, Plate 1, Well A03, 458012
 [9824D314035DA6EA] Jun 25, 2025 11:21:00 Pressure = 62 bar (61 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.392	F 0.47	2.2%	16120	
2	0.490	F 0.59	2.1%	14905	
3	0.687	F 0.82	11.5%	83643	Acetylated F peak
4	0.833	F 1.00	49.1%	356104	Consistent with F
5	0.992	F 1.19	3.5%	25635	
6	1.227	A 1.01	30.9%	224199	A peak - REVIEW
7	1.378	A 1.14	0.7%	4928	
Total Area:				725534	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range