


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2061, Rack A, Plate 1, Well F04, 490390
 [DC238E05F637D99F] Jan 30, 2026 12:12:56 Pressure = 89 bar (88 to 91)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.417	F 0.49	1.0%	16391	
2	0.503	F 0.60	0.3%	4458	
3	0.697	F 0.82	16.6%	271436	Acetylated F peak
4	0.828	F 0.98	73.5%	1200709	Consistent with F
5	1.200	A 1.00	8.6%	140180	A peak
Total Area:				1633174	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range