


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1359, Rack A, Plate 1, Well D01, 27431
[999CF51274D2426A] Apr 03, 2024 10:47:48 Pressure = 61 bar (61 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.487	F 0.55	6.3%	72670	
2	0.777	F 0.88	14.2%	164369	Acetylated F peak
3	0.888	F 1.01	65.1%	751684	Consistent with F
4	1.247	A 0.99	14.4%	166369	A peak
Total Area: 1155092					

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range