


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1901, Rack A, Plate 1, Well C05, 169958
[D9356B31714CF3BD] Sep 03, 2025 11:50:53 Pressure = 61 bar (61 to 62)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.46	1.4%	11625	
2	0.488	F 0.56	0.7%	6172	
3	0.693	F 0.80	12.8%	109550	Acetylated F peak
4	0.842	F 0.97	71.1%	606185	Consistent with F
5	1.215	A 1.01	14.0%	119620	A peak
Total Area:				853152	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range