


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2133, Rack A, Plate 1, Well F06, 173202  
[9C735BB96FEBF315] Apr 29, 2026 13:24:35 Pressure = 66 bar (65 to 67)

AF



| PEAK        | RT    | REL RT | % CONC | AREA   | COMMENT           |
|-------------|-------|--------|--------|--------|-------------------|
| 1           | 0.412 | F 0.47 | 3.3%   | 15277  |                   |
| 2           | 0.493 | F 0.56 | 3.9%   | 17930  |                   |
| 3           | 0.707 | F 0.81 | 5.4%   | 24704  | Acetylated F peak |
| 4           | 0.863 | F 0.98 | 25.1%  | 115402 | Consistent with F |
| 5           | 1.007 | F 1.15 | 5.4%   | 24959  | 3                 |
| 6           | 1.237 | A 1.01 | 55.9%  | 256671 | A peak - REVIEW   |
| 7           | 1.397 | A 1.14 | 0.9%   | 4089   |                   |
| Total Area: |       |        |        | 459032 |                   |

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range