


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2197, Rack A, Plate 1, Well C03, 487115
[DC3C86CAEAAAEC62] Jul 15, 2026 13:09:26 Pressure = 84 bar (83 to 86)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.525	F	0.63	1.0%	15592	
2	0.687	F	0.83	17.5%	267041	Acetylated F peak
3	0.823	F	0.99	74.5%	1139243	Consistent with F
4	1.160	A	1.00	7.0%	106916	A peak
Total Area: 1528792						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range