


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2173, Rack A, Plate 1, Well F04, 493870  
[DC2CA862FB386E62] Jun 10, 2026 15:12:54 Pressure = 66 bar (66 to 68)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.510	F 0.60	1.9%	26448	
2	0.697	F 0.82	12.8%	176480	Acetylated F peak
3	0.848	F 1.00	60.2%	830903	Consistent with F
4	1.195	A 1.00	25.2%	347412	A peak - REVIEW
Total Area:				1381243	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range