


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2143, Rack A, Plate 1, Well B12, 503312
[8C271E1C1C9EA4E8] May 08, 2026 12:10:21 Pressure = 64 bar (63 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.510	F 0.60	1.7%	22998	
2	0.713	F 0.84	11.8%	157961	Acetylated F peak
3	0.853	F 1.00	54.2%	725275	Consistent with F
4	1.213	A 1.01	32.3%	433054	A peak - REVIEW
Total Area: 1339288					

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range