


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1893, Rack A, Plate 1, Well A08, 830810
 [DC24385E49B9A442] Aug 27, 2025 10:02:10 Pressure = 63 bar (62 to 64)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.405	F 0.48	1.5%	11260	
2	0.488	F 0.58	1.3%	10167	
3	0.677	F 0.81	14.9%	112624	Acetylated F peak
4	0.835	F 0.99	59.1%	445383	Consistent with F
5	0.983	F 1.17	2.1%	15683	
6	1.213	A 1.01	21.0%	158221	A peak - REVIEW
Total Area:				753338	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range