


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2192, Rack A, Plate 1, Well H10, 486241
[CD28D110098A46C2] Jul 06, 2026 16:03:38 Pressure = 80 bar (77 to 81)

FAS



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.433	F	0.24	2.1%	90703	
2	0.547	F	0.30	3.2%	133643	
3	0.860	F	0.47	25.8%	1086795	
4	1.638	F	0.90	30.3%	1277211	
5	1.783	F	0.98	25.6%	1081521	Consistent with F
6	2.625	F	1.44	1.1%	47444	
7	3.503	A	0.99	6.5%	275571	A peak
8	5.113	S	1.01	5.4%	226095	Consistent with S
Total Area: 4818983						A/V=1.20

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range