


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1982, Rack A, Plate 1, Well E10, 482073
 [8873E307A1F02E60] Nov 19, 2025 12:32:55 Pressure = 68 bar (67 to 70)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.392	F 0.45	1.4%	15859	
2	0.495	F 0.57	0.6%	7351	
3	0.708	F 0.82	12.6%	145672	Acetylated F peak
4	0.852	F 0.99	53.9%	622245	Consistent with F
5	1.012	F 1.17	1.9%	21624	
6	1.240	A 1.01	29.6%	342445	A peak - REVIEW
Total Area:				1155196	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range