


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2038, Rack A, Plate 1, Well A06, 171729-P
 [8936C9F65242A448] Jan 09, 2026 10:46:57 Pressure = 80 bar (79 to 82)

AF



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.390	F 0.47	4.5%	24736	
2	0.487	F 0.58	2.5%	13421	
3	0.718	F 0.86	10.7%	58203	
4	0.847	F 1.01	27.6%	150611	Consistent with F
5	1.017	A 0.86	4.8%	26456	
6	1.215	A 1.02	49.1%	268302	A peak - REVIEW
7	1.395	S 0.90	0.8%	4412	
Total Area:				546141	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range