


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2200, Rack A, Plate 1, Well F09, 500810
[9868D72DC60ACC62] Jul 17, 2026 16:37:26 Pressure = 82 bar (82 to 85)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.682	F	0.82	19.8%	254393	Acetylated F peak
2	0.817	F	0.98	72.2%	926733	Consistent with F
3	1.153	A	1.01	7.9%	101730	A peak
Total Area: 1282856						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range