


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2077, Rack A, Plate 1, Well B12, 492104  
 [DC366792FDD1D39F] Feb 18, 2026 10:32:56 Pressure = 57 bar (56 to 58)

FA



| PEAK        | RT    | REL RT | % CONC | AREA    | COMMENT           |
|-------------|-------|--------|--------|---------|-------------------|
| 1           | 0.415 | F 0.48 | 0.9%   | 13343   |                   |
| 2           | 0.517 | F 0.59 | 0.7%   | 10716   |                   |
| 3           | 0.727 | F 0.83 | 16.8%  | 260593  | Acetylated F peak |
| 4           | 0.862 | F 0.99 | 56.3%  | 871586  | Consistent with F |
| 5           | 1.027 | F 1.18 | 2.3%   | 36073   |                   |
| 6           | 1.258 | A 1.01 | 23.0%  | 355566  | A peak - REVIEW   |
| Total Area: |       |        |        | 1547877 |                   |

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range