


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 1850, Rack A, Plate 1, Well A06, 785274  
[D97099D32842D917] Jul 02, 2025 10:38:58 Pressure = 61 bar (61 to 62)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.403	F 0.48	1.8%	10639	
2	0.492	F 0.58	3.4%	20461	
3	0.702	F 0.83	15.0%	90292	Acetylated F peak
4	0.873	F 1.04	66.8%	401820	Consistent with F
5	1.223	A 1.01	13.0%	78225	A peak
Total Area:				601437	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range