


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2133, Rack A, Plate 1, Well G10, 503306
[99220FEC383E2460] Apr 29, 2026 13:56:55 Pressure = 65 bar (65 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.518	F 0.59	1.7%	28922	
2	0.717	F 0.82	18.3%	314155	Acetylated F peak
3	0.862	F 0.98	75.0%	1289142	Consistent with F
4	1.222	A 1.00	5.1%	87082	A peak
Total Area:				1719301	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range