


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2176, Rack A, Plate 1, Well D02, 493885
[9969EC417EC7313D] Jun 12, 2026 13:05:41 Pressure = 68 bar (68 to 70)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.517	F 0.61	0.7%	11082	
2	0.700	F 0.83	14.8%	220884	Acetylated F peak
3	0.827	F 0.98	69.9%	1040671	Consistent with F
4	1.190	A 1.00	14.5%	215710	A peak
Total Area:				1488347	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range