


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2146, Rack A, Plate 1, Well G06, 501867
[C968A190AAF04CE0] May 13, 2026 14:21:03 Pressure = 66 bar (65 to 67)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.513	F 0.59	2.3%	23321	
2	0.708	F 0.81	13.2%	136642	Acetylated F peak
3	0.868	F 0.99	66.2%	684244	Consistent with F
4	1.223	A 1.01	18.3%	189007	A peak
Total Area:				1033214	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range