


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2149, Rack A, Plate 1, Well G05, 502232  
[9979F4494BA51917] May 20, 2026 14:44:18 Pressure = 64 bar (64 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.517	F 0.60	2.8%	33287	
2	0.722	F 0.83	16.5%	192929	Acetylated F peak
3	0.867	F 1.00	65.7%	769602	Consistent with F
4	1.105	A 0.92	0.3%	3400	
5	1.208	A 1.00	14.7%	171718	A peak

Total Area: 1170936

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range