


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1827, Rack A, Plate 1, Well A02, 456682
[C865D522DBA0D9BF] Jun 11, 2025 09:15:56 Pressure = 61 bar (60 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.47	1.7%	9572	
2	0.493	F 0.58	1.4%	7560	
3	0.717	F 0.84	17.2%	94341	Acetylated F peak
4	0.847	F 1.00	72.3%	396516	Consistent with F
5	1.225	A 1.01	7.4%	40627	A peak
Total Area:				548616	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range