


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1380, Rack A, Plate 1, Well B05, 65080
[9D8DB31BB40D3F97] Apr 17, 2024 10:56:58

Pressure = 83 bar (83 to 83)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.392	F 0.46	1.7%	13205	
2	0.478	F 0.56	1.9%	14466	
3	0.743	F 0.88	13.4%	101252	Acetylated F peak
4	0.862	F 1.02	51.9%	392173	Consistent with F
5	0.977	F 1.15	0.9%	6842	
6	1.222	A 1.02	30.2%	228269	A peak - REVIEW
Total Area:				756207	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range