


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2146, Rack A, Plate 1, Well A12, 503315
[DC79A4D5D7A7B39D] May 13, 2026 12:03:37 Pressure = 66 bar (66 to 68)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.515	F 0.60	2.7%	23483	
2	0.707	F 0.83	17.3%	152952	Acetylated F peak
3	0.855	F 1.00	67.6%	596899	Consistent with F
4	1.222	A 1.01	12.4%	109032	A peak

Total Area: 882366

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range