


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2113, Rack A, Plate 1, Well H09, 502304
[DC2242932889799D] Apr 07, 2026 14:22:01 Pressure = 59 bar (58 to 60)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.522	F 0.58	1.6%	24540	
2	0.723	F 0.80	13.9%	207556	Acetylated F peak
3	0.863	F 0.95	67.9%	1010827	Consistent with F
4	1.257	A 1.00	16.6%	246818	A peak
Total Area:				1489741	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range